



Certificate of Conformity No 116-03112

Translation

| | |
|----------------------------|---|
| <i>Object</i> | Luminaire for medicinal purpose |
| <i>Type</i> | CHROM-NT, serial number 12160051 / CH/100/NT Dimensions (L x W x H) : 1537 mm x 635 mm x 110 mm |
| <i>Ballast</i> | Integrated |
| <i>Applicant</i> | ZENIUM S.A. Rue Edouard Michelin 100 F-54 710 Fléville Devant Nancy |
| <i>Requirements</i> | Conformity according IEC 62471:2006 "Photobiological safety of lamps and lamp systems" |
| <i>Confirmation</i> | According to the Measurement Report METAS Nr. 116-03110 from 18 th of November ²⁰¹⁶ the luminaire CHROM-NT is classified in the exempt group. It can be considered as save concerning of photo-biological risks. A lesion of the retina caused by blue light can be excluded. |
| <i>Date of Examination</i> | 17.11.2016 |

3003 Bern-Wabern, 19th of January 2017

For the Examination

Beat Imhof

Sector Length, Optics and Time

Dr Rudolf Thalmann, Head of Sector