



Certificate of Conformity No 116-03260


Translation

<i>Object</i>	Luminaire for medicinal purpose
<i>Type</i>	NANO, serial number 13160051 / NA/150/LD Dimensions (L x W x H) : 1280 mm x 416 mm x 54 mm
<i>Ballast</i>	Integrated
<i>Applicant</i>	ZENIUM S.A. Rue Edouard Michelin 100 F-54 710 Fléville Devant Nancy
<i>Requirements</i>	Conformity according IEC 62471:2006 <i>Photobiological safety of lamps and lamp systems</i>
<i>Confirmation</i>	According to the Measurement Report METAS Nr. 116-03245 from 2 nd of March 2017 the luminaire NANO is classified in the exempt group. It can be considered as safe concerning of photo biological risks. A lesion of the retina caused by blue light can be excluded.
<i>Date of Examination</i>	01.03.2017

3003 Bern-Wabern, 2nd of March 2017

For the Examination

Sector Length, Optics and Time


Beat Imhof


Dr Rudolf Thalman, Head of Sector